

To whom it may concern;

## Calibration Certificate

Manufacturer : EKO Instruments Co., Ltd.  
Model : MR-60  
Description : 4-Component Radiometer

Calibration date : May 12, 2016  
Serial number : S15115.07  
EKO Ref. number EX16-044

We hereby certify that calibration results of above instrument are as follows;

		Sort-wave Radiation	Long-wave Radiation	Unit
Sensitivity	(Upper)	6.95	3.08	$\mu\text{V} / \text{W} \cdot \text{m}^{-2}$
	(Lower)	6.86	3.22	$\mu\text{V} / \text{W} \cdot \text{m}^{-2}$
Internal Resistance	(Upper)	66	97	$\Omega$
	(Lower)	64	97	$\Omega$



I. Sunaga / Manager of Mfg. Dept.



T. Konuma / Inspector